



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY

CINCINNATI, OHIO 45268

Office of Ground Water and Drinking Water

Technical Support Center

September 20, 2012

EMSL Analytical, Inc.
200 Route 130 North
Cinnaminson, NJ 08077

Dear Shannon Kauffman,

Based on your application and successful participation in a Unregulated Contaminant Monitoring Rule 3 (UCMR 3) Proficiency Testing (PT) Study for each of the listed methods, EPA has given the status of "approved" to your laboratory for the method(s) listed below. The application and PT criteria are listed in the "UCMR 3 Laboratory Approval Requirements and Information Document, Version 2.0." Your laboratory is now "approved" to conduct UCMR 3 analyses using the following method(s):

LabID: NJ00337

Method Name

Inorganic anions by IC (Chlorate)

EPA 300.1

Approval Date

9/5/2012

End of Method List

The information will be forwarded to the UCMR 3 Webmaster for inclusion in the list of UCMR 3 approved laboratories. Your approval status will be maintained during UCMR 3 by continuing to meet the criteria given in the "UCMR 3 Laboratory Approval Requirements and Information Document, Version 2.0," and any revisions to the aforementioned document. Please be aware that you are only permitted to conduct UCMR 3 analyses using those methods for which you have EPA approval. Should you wish to comment on any of these determinations, please write to:

UCMR 3 Laboratory Approval Coordinator
USEPA, Technical Support Center
26 W. Martin Luther King Drive (MS 140)
Cincinnati, OH 45268
UCMR_Sampling_Coordinator@epa.gov

Sincerely,

Michella S. Karapondo
UCMR Laboratory Approval Coordinator



UCMR 3 Proficiency Testing Study 4 Results

September 19, 2012

EMSL Analytical, Inc.

Lab ID: NJ00337

PT Study ID: UCMR_PT04

Study Dates: 8/21/2012 - 9/5/2012

MethodID: EPA_300_1 PT Method Results: PASS

Analyte Name	Analyte ID	Unit	Your PT Result	True Value	Acceptance Limits	PT Analyte Result
chlorate	1007	µg/L	45.7	47.0	28.0 - 66.0	PASS